



ಸಿಧ್ಧಗಂಗಾ ವೈದ್ಯಕೀಯ ಮಹಾವಿದ್ಯಾಲಯ ಮತ್ತು ಸಂಶೋಧನಾ ಸಂಸ್ಥೆ SIDDAGANGA MEDICAL COLLEGE & RESEARCH INSTITUTE

(A unit of Sree Siddaganga Math)

Approved by National Medical Commission, New Delhi, and Government of Karnataka
Affiliated to Rajiv Gandhi University of Health Sciences, Bangalore.

Ref. No.: SMCRI/BOQ/26-27/

Date: 11/05/2026

Last Date & Time for Receipt of Quotation: 30/5/26 up to 14:00 Hours

Date & Time of Opening of Quotation: 9/5/26 at 16:00 Hours

Details are also available on www.smcri.edu.in → Procurement Tab

INVITATION FOR QUOTATIONS - Supply of EUS system

Dear Sir/Madam,

Sub: Invitation for Quotations – Supply of EUS system

You are invited to submit your **most competitive quotation** for the supply of the EUS system as detailed below.

1. Schedule of the items - BOQ

SI NO	DESCRIPTION	QTY
1	<p>Endoscopic Ultrasound (EUS) System Complete high-definition Endoscopic Ultrasound system with Video Processor, Light Source, Ultrasound Processor, Curvilinear Echoendoscope, Medical Grade Monitor, and Documentation System.</p> <p>Video Processor and Light Source Latest generation processor compatible with high-definition endoscopes.</p> <ul style="list-style-type: none">• Full HD or 4K image quality.• LED or Xenon light source with long life.• Image enhancement technologies• USB and network connectivity. <p>Ultrasound Processor</p> <ul style="list-style-type: none">• B-mode imaging• Color Doppler• Power Doppler• Frequency range: 5–12 MHz or better.• Adjustable gain, focus, depth, and dynamic range.• Digital image storage and export. <p>Curvilinear Echoendoscope</p> <ul style="list-style-type: none">• Electronic curved linear array.• Forward oblique viewing.• Compatible with FNA/FNB needles.• Working channel: minimum 3.7 mm.• Scanning angle: minimum 180°.• Frequency range: 5–10 MHz or better. <p>Bending angles:</p> <ul style="list-style-type: none">• Up ≥130°• Down ≥90°• Right/Left ≥90° <p>Water jet facility preferred.</p> <p>Video Monitor</p> <ul style="list-style-type: none">• Medical-grade monitor.• Minimum 27-inch Full HD; 4K preferred.• HDMI/DVI inputs.• Height-adjustable stand.	1No

	<p>Accessories to be Supplied</p> <ul style="list-style-type: none"> • Leak tester. • Water bottle and tubing. • Suction valves and biopsy valves. • Cleaning adapters. • Scope storage case. • User manuals. • Foot switch <p>Consumables</p> <ul style="list-style-type: none"> • Minimum 10 EUS FNA needles (22G). • Minimum 5 EUS FNB needles. • Cytology brush if applicable. • Sterile water bottles and tubing sets. 	
--	--	--

2. Submission Mode

- a) The procurement shall be for the full quantity as described above. Quotes shall be submitted to the secure email id only – **purchase.smcri@gmail.com**
- b) **Offline / physical submission is not permitted** and will be summarily rejected.

3. BOQ Format

- a) The BOQ is provided in **Schedule of Items (BOQ)**.
- b) Modification of BOQ structure, columns, rows, formulas, or item descriptions is strictly prohibited.

4. Quotation

- a) Vendors shall quote the **unit rate and total amount** against each BOQ item.
- b) Rates must be quoted in **Indian Rupees (INR)** only.
- c) Quotation in foreign currency is **not permitted**.

5. Evaluation of Quotations

- a) The Purchaser will evaluate and compare quotations that are **substantially responsive**, properly signed, and conforming to the specifications and terms.

6. Award of Purchase Order

- a) The vendors will be invited for the Purchase Committee Meeting for technical and financial discussions.
- b) The Purchaser reserves the right to **accept or reject any or all quotations** and cancel the process at any time before award.

7. Taxes & Duties

- a) Quoted rates shall be **inclusive / exclusive of GST** and other applicable taxes, as specified.
- b) GST shall be quoted strictly as per **prevailing Government rules**.

8. Validity of Quote

The financial quote shall remain valid for a minimum period of **90 days** from the date of submission.

9. Rejection of Quotation

The financial Quote is liable to be rejected if:

- a) BOQ format is altered
- b) Rates are not filled for any item
- c) Multiple rates are quoted for a single item
- d) The quote should be strictly sent only to the secure mail id without any CC or BCC.

10. Final Acceptance

The purchaser reserves the right to **accept or reject any or all bids** without assigning any reason thereof.

11. General Conditions

- a) The Quotations should be addressed to "The Secretary – Medical Services, SMCRI, Tumkur "(Should not be addressed to any other personnel of the organisation)
- b) For technical Specification regarding procurement, please contact: HOE, SMCRI. Email: hoe@smcri.edu.in, Phone: 8880340730
- c) **Delivery period, payment terms, validity of quotation, and warranty period** shall be clearly mentioned. The warranty period for the above-mentioned items should be for a minimum of 5 years.
- d) The supplied items shall be **new, unused, and of approved quality**.

We look forward to receive your quotations at the earliest.

SIDDAGANGA HOSPITAL
A Unit of Sree Siddaganga Math-SMCRI



CEO